




# OTE / BTE (On The Ear/ Behind The Ear)


## EVOK™ 917 SPECIFICATIONS

Ear Simulator [IEC-118-0]*	Thin Tube / Ear Hook	2cc Coupler [ANSI S3.22-1996]*	Thin Tube / Ear Hook
OSPL90 Maximum	128/133 dB SPL	OSPL90 Maximum	120/125 dB SPL
OSPL90 @ 1600 Hz	118/125 dB SPL	HF Average OSPL90	110/117 dB SPL
Peak Gain	50/60 dB	Peak Gain	42/52 dB
Full-On-Gain @ 1600 Hz	43/52 dB	HFA Full-On-Gain	35/44 dB
R.T.G @ 1600 Hz	41/48 dB	Reference Test Gain	33/40 dB
Frequency Range: f1-f2	200-8000 Hz	Frequency Range: f1-f2	200-8000 Hz
THD at 500 Hz	< 3%	THD at 500 Hz	< 3%
800 Hz	< 3%	800 Hz	< 3%
1600 Hz	< 3%	1600 Hz	< 3%
Eq. Input Noise Level	< 26 dB SPL	Eq. Input Noise Level	< 26 dB SPL
Attack Time	6 mS @ 2kHz	Attack Time	6 mS @ 2kHz
Release Time	30 mS @ 2kHz	Release Time	30 mS @ 2kHz
Battery Current Drain	0.95 mA	Battery Current Drain	0.95 mA
Battery Life [10A]	84 hrs	Battery Life [10A]	84 hrs

\*Test Settings Provided Upon Request



- ▶▶ Speech-Pro™ Auto Fit (Optional)
- ▶▶ Data Logging
- ▶▶ Speech Mapping Compatible w/ Graphic Equalizer Controls
- ▶▶ Focus-Mic System™
- ▶▶ Feedback Shield™
- ▶▶ Tri-Mode Layered Noise Reduction
- ▶▶ Dynamic Speech Enhancement (DSE)
- ▶▶ 12 Bands, 8 Channels
- ▶▶ 4 Programmable Memories
- ▶▶ Memory Change Indicator w/ Adjustable Level and Frequency
- ▶▶ Low Battery Indicator
- ▶▶ Power on Delay w/ Adjustable Level
- ▶▶ NOAH Compatible



**EVOK™**

PERSONA  
MEDICAL